

# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm  
State of Alaska

APR 20 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm S/N <u>130/80</u>	
A	Name <u>A.W. Anderson</u>	ID# <u>2467</u>	Date <u>4/15/10</u>
	Agency <u>Seldovia Police Dept.</u>	Phone # _____	
	Instrument Location <u>Seldovia Police Dept.</u>		
B	Alco S/N <u>X 301530</u>	Target Value <u>.078</u>	High Pressure <u>500 PSI</u>
	Alco Test Value Average <u>.076</u>	<u>.076</u>	
	1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco	
Signature <u>AW Anderson</u>		CAB 4/26/10	
(OVER)			

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:



- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 26<sup>th</sup> day of April, 2010.

Carolyn M. Noland (Notary Seal Stamp)

Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



of Datamaster cdm  
State of Alaska  
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130180

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130180

APRIL 15, 2010

OPERATOR'S NAME:

ANDERSON AM

OPERATOR'S NUMBER: 2467

SUBJECT'S LAST NAME:

SIXTY DAY CALIBRATION

SUBJECT'S FIRST NAME/MI :

N/A

D.L. #: 1234567890

DEPT/AGENCY: SOU1

CASE/REPORT: N/A

TEST TYPE: U

ALCO TARGET VALUE: .078

ALCO S/N: X301530

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 30.33 in

ALCO TARGET .079 14:55

BLANK TEST .000 14:56

INTERNAL STANDARD VERIFIED 14:56

ALCO TV 30.33 in .076 14:56

BLANK TEST .000 14:57

SUBJECT SAMPLE .000 14:58

BLANK TEST .000 14:58

ALCO TV 30.33 in .076 14:59

BLANK TEST .000 15:00

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130180

APRIL 15, 2010

TIME 14:52

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 30.35 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!""#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHI  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abdefghijklmno  
pqrstuvwxyz{|}~